

The department of ENT, KEM hospital, invites quotations for equipments to set up Advanced ENT Endoscopy Centre

The department of ENT, KEM hospital, Mumbai invites quotations for equipments. Quotations should be addressed to DJST, Seth G.S. Medical College and KEM Hospital Mumbai. Quotations should be sent in a sealed envelope, the address being

Diamond Jubilee Society trust,
1st Floor, College building,
KEM Hospital,
Mumbai 400012

- Specifications for Rhino-Pharyngo-Laryngo-Fiberscope Direction of view: 0° Angle of view: 70° Working length: 30 cm O.D.: 3.5 mm Following accessories to be included: Case, Pressure Compensation Cap, Leakage Tester
- Specifications for Endoscopic video unit for use with one-chip camera heads and video endoscopes, incl. LED light source similar to Xenon technology, with integrated digital Image Processing Module, 15-16" LCD TFT monitor with LED backlight, USB/SD memory module, color systems PAL/NTSC, power supply 100 - 240 VAC, 50/60 Hz including: US USB Silicone Keyboard, with touchpad, US character set USB Flash Drive, 30-35 GB 400 F Mains Cord, length 300 cm
- Specifications for One-Chip Camera Head, color system PAL, soakable, gas-sterilizable, with integrated Parfocal Zoom Lens, f = 25 - 50 mm (2x), 2 freely programmable camera head buttons
- Specifications for Fiber Optic Light Cable, with straight connector, diameter 3.5 mm, length 230 cm
- Specifications for Stroboscopy Kit, for use with Camera system consisting of: Footswitch and Microphone Set.
- Specifications for Strobo-Laryngoscope with integrated lateral telescope 70°, oval sheath, 7.2 x 9.3 mm, working length 17 cm, autoclavable, fiber optic light transmission incorporated
- Specifications for Rhino-Pharyngo- Laryngo-Fiberscope Direction of view: 0° Angle of view: 70° Working length: 30 cm O.D.: 3.5 mm Following accessories are included: Case, Pressure Compensation Cap, Leakage Tester
- Specifications for Wide Angle Lateral Telescope 70°, enlarged view, diameter 4 mm, length 18 cm, autoclavable. Fiber optic light transmission incorporated.
- Specifications for Specifications for Tele-Otoscope, with Straight Forward Telescope 0°, diameter 3 mm, length 6 cm, autoclavable, fiber optic light transmission incorporated
- Specifications for Tele-Otoscope, with Forward-Oblique Telescope 30°, diameter 3 mm, length 6 cm, autoclavable, fiber optic light transmission incorporated

The department of ENT of KEM hospital invites quotations (Last date 24/12/2016).

- The Quotation should be addressed to "Diamond jubilee society trust", Seth GS Medical college and KEM hospital, Mumbai.
- You are requested to submit the quotation to following address –
- Mrs. K. D. Angane ,
Department of ENT,
107 ,ENT OPD , 1st floor, old building,
KEM Hospital campus, Mumbai, Tel No 02224107307
- **For further query please contact**
Dr. Hetal Marfatia (drhetalent@gmail.com , 02224107012) o
Dr. Arpit Sharma (drarpitsharma@gmail.com)

